




MYTHIC CRP Calibrators

CRP CALIBRATORS

CALIBRATOR

LOT 2024032101

 2025-03-20

ASSAY VALUES

Instrument	Level (multi-point)	Reference Value (mg/L)
Orphée Mythic 3CRP	Calibrator 1	5.8
	Calibrator 2	20.7
	Calibrator 3	55.1
	Calibrator 4	108.5
	Calibrator 5	147.6
	Calibrator 6	176.0

NOTE Before using, refer to the instruction sheet for mixing directions.
All brands and products are trademarks or registered trademarks of their respective companies.

ORPHÉE SA
19 chemin du Champ-des-Filles
CH-1228 Plan-les-Ouates / Geneva
Switzerland

EC REP

PZ CORMAY S.A.
ul. Wiosenna 22,
05-092 Łomianki, POLSKA

CE IVD