



MYTHIC CRP Calibrators

CRP CALIBRATORS

CALIBRATOR

LOT 2023121301

 2024-12-12

ASSAY VALUES

Instrument	Level (multi-point)	Reference Value (mg/L)
Orphée Mythic 3CRP	Calibrator 1	5.0
	Calibrator 2	19.5
	Calibrator 3	52.7
	Calibrator 4	113.7
	Calibrator 5	162.5
	Calibrator 6	214.6

NOTE Before using, refer to the instruction sheet for mixing directions.
All brands and products are trademarks or registered trademarks of their respective companies.

ORPHÉE SA
19 chemin du Champ-des-Filles
CH-1228 Plan-les-Ouates / Geneva
Switzerland

EC REP

PZ CORMAY S.A.
ul. Wiosenna 22,
05-092 Łomianki, POLSKA

CE **IVD**