

ADVISORY NOTE

Date of issue	29th March 2023	
Note number	ND_2303_05	
Product name and catalogue numer	Name	Cat. No
	CORMAY CHOLINESTERASE 30	2-057
	CORMAY CHOLINESTERASE 60	2-058
	CORMAY CHOLINESTERASE 120	2-059
	ACCENT-200 CHOLINESTERASE	7-256
	ACCENT-300 CHOLINESTERASE	7-356
	A-400 CHOLINESTERASE	7-456
	PRESTIGE 24i CHOLINESTERASE	4-256
	PRESTIGE 24i CHOLINESTERASE	4-456
	HC-CHOLINESTERASE	4-596
	OS-CHOLINESTERASE	9-458
	B50-CHOLINESTERASE	5-533
	A-800 CHOLINESTERASE	7-838
	EQ CHOLINESTERASE	B02-R0046
Lot number	all lots	
Expiry date	n/a	
Issue description		
<p>Results obtained during routine reagent quality control exceeded the internal acceptance criteria. The values assigned for cholinesterase in:</p> <ul style="list-style-type: none"> - CORMAY SERUM HN (Cat. No 5-172), lot HN10, - CORMAY SERUM HP (Cat. No 5-173), lot HP10 <p>were re-determined and reported in the erratum NDE_2303_04 (Appendix 1).</p>		
Necessary actions		
<ul style="list-style-type: none"> • Read carefully Advisory Note and the erratum NDE_2303_04 (Appendix 1) and use the re-determined values of controls during determinations of cholinesterase. • Perform quality control prior to ordering patient samples. • Please complete and return the verification form. <p>The note is only for reference. Do not discard the product!</p>		

Please provide this information to all Users of the above product and keep the Advisory Note as a quality system documentation.

Please complete and return the attached Verification Form to confirm receipt of the Advisory Note.

We apologize for any inconvenience caused by the above situation.

Sincerely,