

**ADVISORY NOTE**

Date of issue	29 <sup>th</sup> March 2023	
Note number	ND_2303_06	
Product name and catalogue number	<b>Name</b> URIC ACID STANDARD 5	<b>Cat. No</b> 5-125
Lot number Expiry date	<b>Lot number</b> XXX-66 XXX-67	<b>Expiry date</b> 2024-07 2025-04
Issue description		
Results obtained during routine reagent quality control exceeded the internal acceptance criteria. To solve the problem the value of the <b>URIC ACID STANDARD 5</b> (Cat. No. 5-125) was redetermined, based on the primary standard NIST 1950. New standard value is:		
<b>REAGENT</b>	<b>LOT NUMBER</b>	<b>STANDARD VALUE</b>
URIC ACID STANDARD 5	XXX-66	4.70 mg/dl (279.56 µmol/l)
	XXX-67	4.74 mg/dl (281.94 µmol/l)
Necessary actions		
<ul style="list-style-type: none"> <li>• Read carefully Advisory Note and <b>use the re-determined values</b> of standard during determinations of uric acid.</li> <li>• Perform <b>calibration and quality control</b> prior to ordering patient samples.</li> <li>• Complete and return the Verification Form.</li> </ul> <p>The note is only for reference. <b>Do not discard the product!</b></p>		

Please provide this information to all Users of the above product and keep the Advisory Note as a quality system documentation.

Please complete and return the attached Verification Form to confirm receipt of the Advisory Note.

We apologize for any inconvenience caused by the above situation.

Sincerely,

Quality Assurance Coordinator