

**ADVISORY NOTE**

Date of issue	3 <sup>rd</sup> April 2023
Note number	ND_2304_01
Product name and catalogue numer	CORMAY RPR CARBON 500 (Cat. No 6-256)
Lot number	211-57
Expiry date	2024-11-30
Issue description	
<p><b>RPR-Control (+) (Lot 11-57)</b> which is included in CORMAY RPR CARBON 500 kit (Cat. No: 6-256) will be withdrawn from the market because of the noticed problem with absence of agglutination.  <b>This problem concerns <u>not entire kit, but only the RPR-Control (+)</u>.</b></p> <p>Please contact with our Area Sales Manager or Customer Service for exchange of the RPR-Control (+) with above mentioned lot number and delivery of new ones. The exchange conditions will be agreed individually with customers.</p>	
Necessary actions	
<ul style="list-style-type: none"> <li>• Read carefully the Advisory Note ND_2304_01.</li> <li>• Inform end users of the need to <b>discontinue use of the RPR-Control (+) (Lot 11-57)</b></li> <li>• Please contact with our Area Sales Manager or Customer Service and inform us about the stock levels.</li> <li>• Please complete and return the verification form.</li> </ul> <p>The note is only for reference. <b>Do not discard the product!</b></p>	

**Please provide this information to all Users of the above product and keep the Advisory Note as a quality system documentation.**

**Please complete and return the attached Verification Form to confirm receipt of the Advisory Note.**

**We apologize for any inconvenience caused by the above situation.**

**Sincerely,**

Quality Assurance Coordinator