

ADVISORY NOTE

Date of issue	28.08.2023
Note number	ND_2308_01
Product name and catalogue numer	CORMAY MULTICALIBRATOR LEVEL 2 Cat. No 5-175, 5-177
Lot number	307-06
Expiry date	2025-09
Issue description	
<p>In the value sheet for CORMAY MULTICALIBRATOR LEVEL 2 the value for aspartate aminotransferase was incorrectly provided.</p> <p>The correct value has been placed in value sheet (Appendix 1).</p>	
Necessary actions	
<ul style="list-style-type: none"> • Read carefully Advisory Note, corrected value sheet (Appendix 1) and use the correct values of calibrator during determination of aspartate aminotransferase. • Perform calibration and quality control prior to ordering patient samples. • Complete and return the verification form. <p>The note is only for reference. Do not dispose the product!</p>	

Please provide this information to all Users of the above product and keep the Advisory Note as a quality system documentation.

Please complete and return the attached Verification Form to confirm receipt of the Advisory Note.

We apologize for any inconvenience caused by the above situation.

Sincerely,

Quality Assurance Manager