

ADVISORY NOTE

Date of issue	11.03.2024
Note number	ND_2403_02
Product name and catalogue number	Myt-3D VET (Cat. No MYT303-VET)
Lot number	B0224
Expiry date	2024-05-05
Issue description	
<p>While monitoring the quality of Myt-3D control blood (low and normal levels), a decision was made to redetermine the values for the RDW, LY, MO and GR parameters.</p> <p>The re-determined values for the above parameters are given in the revised assay value sheet (Attachment 1). The values for the other parameters, not listed in the advisory note, have not changed.</p>	
Necessary actions	
<ul style="list-style-type: none"> • Read carefully the attachment and use the re-determined values of RDW, LY, MO, GR parameters for control blood. • Perform quality control prior to ordering patient samples. • Please complete and return the verification form. 	
The note is only for reference. Do not discard the product!	

Please provide this information to all Users of the above product and keep the Advisory Note as a quality system documentation.

Please complete and return the attached Verification Form to confirm receipt of the Advisory Note.

We apologize for any inconvenience caused by the above situation.

Sincerely,

Quality Assurance Manager